

BLOOD MIXER TECHNICAL REQUIREMENTS

1. The instrument has to be intended for mixing of standard test tubes, KS EDTA tubes, reagent bottles and controls.
2. The material of the instrument has to be stainless steel.
3. The device has to work with 220V 50-60 Hz.
4. The device has to work at 30-33 rpm.
5. The power of the motor has to be 4 W.
6. The warranty period has to be two (2) years.
7. The device has to have CE certificate.